



March 30, 2023

Smith & Nephew, Inc.  
Madison Padgett  
Regulatory Affairs Specialist  
7135 Goodlett Farms Parkway  
Cordova, Tennessee 38016

Re: K230555

Trade/Device Name: Cemented Tibia Baseplate no Taper with JRNY Lock

Regulation Number: 21 CFR 888.3560

Regulation Name: Knee Joint Patellofemorotibial Polymer/Metal/Polymer Semi-Constrained Cemented Prosthesis

Regulatory Class: Class II

Product Code: JWH

Dated: February 27, 2023

Received: February 28, 2023

Dear Madison Padgett:

We have reviewed your Section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act) that do not require approval of a premarket approval application (PMA). You may, therefore, market the device, subject to the general controls provisions of the Act. Although this letter refers to your product as a device, please be aware that some cleared products may instead be combination products. The 510(k) Premarket Notification Database located at <https://www.accessdata.fda.gov/scripts/cdrh/cfdocs/cfpmn/pmnmnm.cfm> identifies combination product submissions. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration. Please note: CDRH does not evaluate information related to contract liability warranties. We remind you, however, that device labeling must be truthful and not misleading.

If your device is classified (see above) into either class II (Special Controls) or class III (PMA), it may be subject to additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 898. In addition, FDA may publish further announcements concerning your device in the Federal Register.

Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal statutes and regulations administered by other Federal agencies. You must comply with all the Act's requirements, including, but not limited to: registration and listing (21 CFR Part 807); labeling (21 CFR Part 801); medical device reporting (reporting of medical device-related adverse events) (21 CFR 803) for devices or postmarketing safety reporting (21 CFR 4, Subpart B) for combination products (see <https://www.fda.gov/combination-products/guidance-regulatory-information/postmarketing-safety-reporting-combination-products>); good manufacturing practice requirements as set forth in the quality systems (QS) regulation (21 CFR Part 820) for devices or current good manufacturing practices (21 CFR 4, Subpart A) for combination products; and, if applicable, the electronic product radiation control provisions (Sections 531-542 of the Act); 21 CFR 1000-1050.

Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21 CFR Part 807.97). For questions regarding the reporting of adverse events under the MDR regulation (21 CFR Part 803), please go to <https://www.fda.gov/medical-devices/medical-device-safety/medical-device-reporting-mdr-how-report-medical-device-problems>.

For comprehensive regulatory information about medical devices and radiation-emitting products, including information about labeling regulations, please see Device Advice (<https://www.fda.gov/medical-devices/device-advice-comprehensive-regulatory-assistance>) and CDRH Learn (<https://www.fda.gov/training-and-continuing-education/cdrh-learn>). Additionally, you may contact the Division of Industry and Consumer Education (DICE) to ask a question about a specific regulatory topic. See the DICE website (<https://www.fda.gov/medical-devices/device-advice-comprehensive-regulatory-assistance/contact-us-division-industry-and-consumer-education-dice>) for more information or contact DICE by email ([DICE@fda.hhs.gov](mailto:DICE@fda.hhs.gov)) or phone (1-800-638-2041 or 301-796-7100).

Sincerely,

  
**Limin Sun-S**

Limin Sun, Ph.D.

Assistant Director

DHT6A: Division of Joint

Arthroplasty Devices

OHT6: Office of Orthopedic Devices

Office of Product Evaluation and Quality

Center for Devices and Radiological Health

Enclosure

## Indications for Use

510(k) Number (if known)

K230555

Device Name

Cemented Tibia Baseplate no Taper with JRNY Lock

Indications for Use (Describe)

Total Knee components are indicated for:

- Rheumatoid arthritis
- Post-traumatic arthritis, osteoarthritis, or degenerative arthritis
- Failed osteotomies, unicompartmental replacement, or total knee replacement.
- Posterior stabilized knee systems are designed for use in patients in primary and revision surgery, where the anterior and posterior cruciate ligaments are incompetent and the collateral ligaments remain intact

The Smith & Nephew Cemented Tibia Baseplate no Taper with JRNY Lock are indicated for use with cement only and are single use devices.

Type of Use (Select one or both, as applicable)

Prescription Use (Part 21 CFR 801 Subpart D)

Over-The-Counter Use (21 CFR 801 Subpart C)

### CONTINUE ON A SEPARATE PAGE IF NEEDED.

This section applies only to requirements of the Paperwork Reduction Act of 1995.

**\*DO NOT SEND YOUR COMPLETED FORM TO THE PRA STAFF EMAIL ADDRESS BELOW.\***

The burden time for this collection of information is estimated to average 79 hours per response, including the time to review instructions, search existing data sources, gather and maintain the data needed and complete and review the collection of information. Send comments regarding this burden estimate or any other aspect of this information collection, including suggestions for reducing this burden, to:

Department of Health and Human Services  
Food and Drug Administration  
Office of Chief Information Officer  
Paperwork Reduction Act (PRA) Staff  
PRASStaff@fda.hhs.gov

*"An agency may not conduct or sponsor, and a person is not required to respond to, a collection of information unless it displays a currently valid OMB number."*

**510(k) Summary**  
**Smith & Nephew – Cemented Tibia Baseplate no Taper with JRNY Lock**

**510(k) Summary Submitted by:** Smith & Nephew, Inc.  
Orthopaedic Division  
7135 Goodlett Farms Parkway  
Cordova, Tennessee 38016

**Date of Summary:** March 30, 2023

**Primary Contact Person:** Madison Padgett, Regulatory Affairs Specialist II  
Phone: (901) 456-8789

**Secondary Contact Person** Rose Beifuss, Senior Manager, Regulatory Affairs  
Phone: (385) 253-2551

**Name of Device:** Cemented Tibia Baseplate no Taper with JRNY Lock  
**Common Name:** Cemented Tibial Baseplate Components

**Device Classification Name and Reference:** 21 CFR 888.3560 Knee joint patellofemorotibial polymer/metal/polymer semi-constrained cemented prosthesis

**Device Class:** Class II

**Panel Code:** Orthopaedics/87

**Product Code:** JWH – Knee joint patellofemorotibial polymer/metal/polymer semi-constrained cemented prosthesis

**Predicate Device:** Primary Predicate – K142807 – ANTHEM PS Total Knee System (S.E. 12/22/2014)  
JWH  
Secondary Predicate – K111711 – JOURNEY II BCS Knee System (S.E. 09/16/2011)  
JWH

**510(k) Summary**  
**Smith & Nephew – Cemented Tibia Baseplate no Taper with JRNY Lock**

**Device Description**

The subject devices of this Special 510(k) are the Cemented Tibia Baseplate no Taper with JRNY Lock. The subject Cemented Tibia Baseplate no Taper with JRNY Lock are tibia baseplate components, and a line extension of the ANTHEM Tibia Baseplate cleared under premarket notification K142807 (S.E. 12/22/2014). The subject devices were modified by incorporating the identical JOURNEY II locking mechanism for the JOURNEY II BCS Knee System, cleared under premarket notifications K111711 (S.E. 09/16/2011).

The subject Cemented Tibia Baseplate no Taper with JRNY Lock is designed to be implanted with the use of bone cement and is made of Ti-6Al-4V. The proximal face of the implant includes locking mechanism, dimensions, and size options identical to Smith & Nephew's JOURNEY Nonporous Tibia Baseplate, originally cleared under premarket notification K042515 (S.E. 03/14/2005) and updated under the JOURNEY II BCS Knee System premarket notifications K111711 (S.E. 09/16/2011) respectively. The distal side of the implant has a central stem and two large posterior-directed fins. The subject device includes the same stem length for all sizes, though the fin sizes vary between baseplate sizes. The distal side of the implant has an identical design to Smith & Nephew's ANTHEM Tibia Baseplate, cleared under premarket notification K142807 (S.E. 12/22/2014). The subject Cemented Tibia Baseplate no Taper with JRNY Lock includes multiple sizes in both left and right versions.

The subject inserts of the system, compatible with the subject Cemented Tibia Baseplate no Taper with JRNY Lock, have been previously cleared via premarket notification K220896 (S.E. 04/26/2022) to include the JOURNEY II Locking Mechanism. The articulating surface of the subject inserts have remained identical to the primary predicate system, ANTHEM PS Total Knee System cleared via premarket notification K142807 (S.E. 12/22/2014). Therefore, the existing femoral components cleared for the system via premarket notification K142807 (S.E. 12/22/2014) continue to be compatible and remain unchanged.

**510(k) Summary  
Smith & Nephew – Cemented Tibia Baseplate no Taper with JRNY Lock**

**Indication for Use:**

Total Knee components are indicated for rheumatoid arthritis; post-traumatic arthritis, osteoarthritis, or degenerative arthritis; failed osteotomies, unicompartmental replacement, or total knee replacement. Posterior stabilized knee systems are designed for use in patients in primary and revision surgery, where the anterior and posterior cruciate ligaments are incompetent and the collateral ligaments remain intact.

The Smith & Nephew Cemented Tibia Baseplate no Taper with JRNY Lock are intended for use with cement only and are single use devices.

**Technological Characteristics**

The subject devices were modified by incorporating the identical locking mechanism cleared under the JOURNEY II BCS Knee System premarket notification K111711 (S.E. 09/16/2011) to the primary predicate ANTHEM Tibia Baseplate cleared under K142807 (S.E. 12/22/2014). A review of the technological characteristics indicates that the subject Cemented Tibia Baseplate no Taper with JRNY Lock are equivalent to existing, legally marketed predicate devices with regards to mechanical performance and that there are no new issues related to the safety and effectiveness of the subject devices. A review of this testing has demonstrated that there are no new issues related to the safety and effectiveness of the subject devices. Clinical data was not needed to support the safety and effectiveness of the subject devices.

**Substantial Equivalence Information**

The subject devices are substantially equivalent to the predicate devices listed in the following table in function, intended use, indications for use, design, and material composition.

**Table 5.1: Substantially Equivalent Predicates**

<b>Manufacturer</b>	<b>Submission Name</b>	<b>Submission Number</b>	<b>Clearance Date</b>
Smith & Nephew, Inc.	ANTHEM PS Total Knee System	K142807	12/22/2014
Smith & Nephew, Inc.	JOURNEY BCS Knee System	K111711	09/16/2011

**510(k) Summary**  
**Smith & Nephew – Cemented Tibia Baseplate no Taper with JRNY Lock**

**Performance Testing:**

A review of the leveraged mechanical data indicates that the subject Cemented Tibia Baseplate no Taper with JRNY Lock devices are substantially equivalent to one or more of the previously cleared predicate devices listed in **Table 5.1** above. The Component Interlock Strength testing for the JOURNEY II Locking Mechanism from the JOURNEY II BCS Knee System premarket notification K111711 (S.E. 09/16/2022) was reviewed to determine the substantial equivalence.

Bacterial endotoxin testing was completed and met the acceptable endotoxin limits as stated in the FDA Guidance, “Submission and Review of Sterility Information in Premarket Notification (510(k)) Submissions for Devices Labeled as Sterile,” “Pyrogen and Endotoxins Testing: Questions and Answers,” and ANSI/AAMI ST72 (FDA Recognition Number 14-541).

**Conclusion**

As previously noted, this Special 510(k) Premarket Notification is being submitted to request clearance for the Cemented Tibia Baseplate no Taper with JRNY Lock. Based on the similarities to the predicate devices and rationale to support substantial equivalence, the subject devices are substantially equivalent to the commercially available predicate devices listed in **Table 5.1**.